

**NATIONAL INSTITUTE OF MENTAL HEALTH AND NEUROSCIENCES
(Institute of National Importance)-Bengaluru-560029**

Ph No.26995024/23/25 /26995780

Email: aaos@nimhans.ac.in

STR-D1/EQPT/LIM.TENDERENQUIRY/638/BP/analgesiameter/2019-20

Date:28.08.2019

Dear Sir,

Sub: Analgesia Meter and Electronic von Frey Anesthesiometer Care Analyzer

With reference to subject cited above, kindly make arrangements to send your Quotation/ Proforma Invoice in a sealed cover to the undersigned for the supply of the following items as detailed below on or before **28.09.2019 (before 04:00 P:M)** positively by **post/ courier/Hand**. Quotation sent by Fax will not be accepted. Please enclose the copies of your Tin Number, Pan Number and GST number.

SI No.	Item Description	Quantity
1	Analgesia Meter and Electronic von Frey Anesthesiometer	01 No
	Downloading of Tender documents from website	www.nimhans.ac.in
	Quotation submission last date	28.09.2019 before 04:00 PM

Note:

1. Sealed quotation in **TWO BID SYSTEM(Financial Bid and Technical Bid in separate cover)** has to be submitted only by manufacturer of the articles or the authorized local supplier of the manufacturer. Copy of authorization letter and related documents to be enclosed with the quotation.
2. Quantity of items may be increased/decreased as per requirement
3. Item wise price, taxes, any surcharges (such as transportation, installation, customs duty, etc.), grand total, warranty, delivery period, contact person details (such as Mobile No., E-mail, local supplier address) should be clearly mentioned and the price mentioned should be till door delivery price.
4. Price quoted, should be valid for **1 year**.
5. Mention our Reference No. and due date on envelope cover.
6. Payment terms: **No advance payment will be made**. Payment will be released only after successful supply in all formats and handing over the Equipment/consumables (whichever applicable) to the End-user.
7. Enclose the copies mandatorily of your GST Number & PAN Number, Vendor Mandate Form(Bank Details)along with a cancelled cheque, without these your quotation will be rejected
8. **Mention the warranty period and delivery period of the item and Technical Compliance Sheet**
9. Quotation must be submitted along with relevant documents counter signed along with seal by the bidder (if applicable such as authorization certificates, Technical Compliance Sheet, Item supplied list of Institutes/Hospitals, catalogues etc.). **If not submitted quotation will be treated as incomplete and may be rejected.**
10. If the purchase of article is approved by the Institute, the article should be supplied/installed within the delivery period quoted or Purchase Order due date else penalty will be levied as per Institute rules.
11. Sealed quotation shall be addressed to **SPECIAL OFFICER(E),PURCHASE SECTION, NIMHANS, HOSUR ROAD, BENGALURU – 560 029**
12. Quotation received after last date and time will not be considered under any circumstances.
13. However, the Institute reserves the right to accept or reject any/all quotations without assigning any reason whatsoever.
14. **For Contact Dr.Bhupesh Mehta,Asst Professor,Dept.of Biophysics,0802699-5878**

**NATIONAL INSTITUTE OF MENTAL HEALTH AND NEUROSCIENCES
(Institute of National Importance)-Bengaluru-560029**

Ph No.26995024/23/25 /26995780

Email: aaos@nimhans.ac.in

STR-D1/EQPT/LIM.TENDERENQUIRY/638/BP/analgesiameter/2019-20

Date:28.08.2019

15. If applicable the Bank Guarantee should be executed for 10% of the Purchase Order value and Agreement to be executed on Rs 200/- stamp paper at the time of Purchase order.

SPECIFICATIONS:

Analgesia Meter and Electronic von Frey Anesthesiometer and its accessories:

Specifications:

- Electronic von Frey Anesthesiometer to assess mechanical allodynia with rigid tips (threshold) and the flexible von Frey hairs (0.5 or 0.8mm diameter).
- The equipment is for use with animals (rats) of weight in range of 1 – 500 to 800 grams.
- LCD Readout: The systems should allow measuring, store and display test readings in grams based upon the amount of pressure applied.
- Resolution of 0.1 gm
- Software should allow export of data in excel format.
- Foot switch.
- Independence from temperature
- Independence from Hygroscoopy
- A cage/mesh to be supplied to place the rats during studies
- Microprocessor Electronics 0.1 Gr Plug-in probes
- Statistical functions – should compute average value and standard deviation for each subject.
- Starter kit with 100 number of 0.8 mm diameter tips.
- Supplier should have good service record in India.
- Warranty 5 years

Yours faithfully

SPECIAL OFFICER(E)