

**NATIONAL INSTITUTE OF MENTAL HEALTH AND NEUROSCIENCES
(Institute of National Importance)-Bengaluru-560029**

Ph No.26995024/23/25 /26995780

Email: aaos@nimhans.ac.in

STR-D1/EQPT/LIM.TENDERENQUIRY/627/BP/Biosafety Cabinet/2019-20

Date:27.08.2019

Dear Sir,

Subject: Quotation for the supply Bio Safety Cabinet

With reference to subject cited above, kindly make arrangements to send your Quotation/ Proforma Invoice in a sealed cover to the undersigned for the supply of the following items as detailed below on or before **26.09.2019 (before 04:00 P:M)** positively by **post/ courier/Hand**. Quotation sent by Fax will not be accepted. Please enclose the copies of your Tin Number, Pan Number and GST number.

SI No.	Item Description	Quantity
1	Bio Safety Cabinet	01 No
	Downloading of Tender documents from website	www.nimhans.ac.in
	Quotation submission last date	27.09.2019 before 04:00 PM

Note:

1. Sealed quotation in **Two Bid System(Technical Bid and Financial Bid in separate covers)** has to be submitted only by manufacturer of the articles or the authorized local supplier of the manufacturer. Copy of authorization letter and related documents to be enclosed with the quotation.
2. Quantity of items may be increased/decreased as per requirement
3. Item wise price, taxes, any surcharges (such as transportation, installation, customs duty, etc.), grand total, warranty, delivery period, contact person details (such as Mobile No., E-mail, local supplier address) should be clearly mentioned and the price mentioned should be till door delivery price.
4. Price quoted, should be valid for **1 year**.
5. Mention our Reference No. and due date on envelope cover.
6. Payment terms: **No advance payment will be made**. Payment will be released only after successful supply in all formats and handing over the Equipment/consumables (whichever applicable) to the End-user.
7. Enclose the copies mandatorily of your GST Number & PAN Number, Vendor Mandate Form(Bank Details)along with a cancelled cheque, without these your quotation will be rejected
8. **Mention the warranty period and delivery period of the item and Technical Compliance Sheet**
9. Quotation must be submitted along with relevant documents counter signed along with seal by the bidder (if applicable such as authorization certificates, Technical Compliance Sheet, Item supplied list of Institutes/Hospitals, catalogues etc.). **If not submitted quotation will be treated as incomplete and may be rejected.**
10. If the purchase of article is approved by the Institute, the article should be supplied/installed within the delivery period quoted or Purchase Order due date else penalty will be levied as per Institute rules.
11. Sealed quotation shall be addressed to **SPECIAL OFFICER(E),PURCHASE SECTION, NIMHANS, HOSUR ROAD, BENGALURU – 560 029**
12. Quotation received after last date and time will not be considered under any circumstances.
13. However, the Institute reserves the right to accept or reject any/all quotations without assigning any reason whatsoever.
14. **For Contact Dr.Padmanabhan,Professor,Dept.of Biophysics,0802699-5101/5102/5901/5878**
15. **If applicable the Bank Guarantee should be executed for 10% of the Purchase Order value and Agreement to be executed on Rs 200/- stamp paper at the time of Purchase order.**

**NATIONAL INSTITUTE OF MENTAL HEALTH AND NEUROSCIENCES
(Institute of National Importance)-Bengaluru-560029**

Ph No.26995024/23/25 /26995780

Email: aaos@nimhans.ac.in

STR-D1/EQPT/LIM.TENDERENQUIRY/627/BP/Biosafety Cabinet/2019-20

Date:27.08.2019

SPECIFICATIONS:

1. Safety Class: BSC Class II Type A2
2. Working width : 1205 mm
3. Cabinet type – 4 Feet. Int. Dims (W x D x H) : 1205 x 526 x 585 mm
Ext. Dims (W x D x H) : 1275 x 792 x 1990 mm (with stand)
4. Air flow type : 70% re-circulation through HEPA filter and 30% exhaust to ambient air through HEPA filter
5. HEPA filter with 99.992~99.995 % efficiency for particles > 0.3um
6. Interior material : Stainless steel 304 polished (AISI 304) and safety glass in side wall
7. Exterior material : SPCC /w epoxy powder coating
8. Tempered safety glass door
9. Supply Voltage : 230V/ 50Hz 1P
10. Integrated digital microprocessor control system with back-light LCD display
11. 9-step adjustable air velocity control
12. UV lamp ON/OFF (30 watt x 1no.)
13. High density Fluorescent lamp ON/OFF (30 watt x 2 nos.).
14. Fan ON/OFF and speed control
15. Adjustable UV sterilization timer
16. Voltage : 220VAC (+/- 10%) 50/60Hz 1P
17. Power Rating : 550 Watt, 3.0 Amp
18. Plug : European std. for 220 VAC
19. Adjustable UV sterilization timer
20. Elapsed time counter and alarm for HEPA filter replacement
21. Stand with Caster
22. Instruction Manual
23. 5 years warranty
24. Supportive documents of installation & functioning of the instrument in premier government institutions.

Yours faithfully

SPECIAL OFFICER(E)

**NATIONAL INSTITUTE OF MENTAL HEALTH AND NEUROSCIENCES
(Institute of National Importance)-Bengaluru-560029**

Ph No.26995024/23/25 /26995780

Email: aaos@nimhans.ac.in

STR-D1/EQPT/LIM.TENDERENQUIRY/627/BP/Biosafety Cabinet/2019-20

Date:27.08.2019