



E-Procurement Global Tender No. STR–D3/NIMHANS/2019-20/IND565 30.12.2019

**GLOBAL TENDER NOTIFICATION
(Through e-procurement portal only)**

The Director, NIMHANS invites tender from eligible tenderers through the Karnataka Government E-Procurement portal for supply of following **EQUIPMENT**.

Sl. No.	Name of the Item	Quantity (in Nos.)	EMD (in ₹)
1.	NEUROSURGICAL OPERATING MICROSCOPE	02	20,00,000/-

Tender Schedule

Downloading of Tender documents from website - https://eproc.karnataka.gov.in/eportal/index.seam	From 30.12.2019 Onwards
Last date for tender enquiry	10.01.2020 upto 11:00 AM
Tender submission last date and time	29.01.2020 upto 11:00 AM
Technical bid will be opened online by the authorized officer on	30.01.2020 at 11:00 AM

Sd/-,
Director

TENDER DOCUMENT
Terms and conditions

1. The bid documents for the above item should be addressed to “The Director, National Institute of Mental Health & Neuro Sciences, Post Box No. 2900, Hosur Road, Bengaluru – 560 029, Karnataka, India” and should be uploaded in E Procurement portal only on or before the due date.
2. The TENDER BID should be valid for **four months** from the due date. **NIMHANS, Bengaluru will not take any responsibility for any technical issues.**
3. **Earnest Money Deposit (EMD):**
 - 3.1. The (EMD) shall be denominated in Indian Rupees [₹] and should be paid in the E-PROCUREMENT PORTAL as per the facility provided.
 - 3.2. The EMD shall not bear any interest and will be refunded to
 - a. successful tenderer on receipt of Agreement and Bank Guarantee.
 - b. unsuccessful tenderer upon finalization of tender bid and award of tender to successful bidder.
 - c. all the tenderers if the tendered item is cancelled or retendered.
4. The tender documents and all correspondence’s relating to the bid should be in **English language only**.



5. Technical bid should comprise of (uploaded copy of documents should be self attested, stamped and better quality – preferably .PDF format) -
- 5.1 Brochure/Catalogue and Data sheet of the equipment.
 - 5.2 Technical Compliance Statement
 - 5.3 Proprietary certificate from the manufacturer mentioning the unique technology or feature/s mentioned apart from the brand name (If applicable).
 - 5.4 Pre requirements required at the installation site (Before submitting the bid, the tenderer should make pre-visit to the installation site and indicate the requirement along with the price bid wherever necessary)
 - 5.5 Delivery Period of the item to be supplied and Time required for installation from the date of purchase order has to be indicated.
 - 5.6 List of Institutes where the equipment has been supplied with copy of purchase orders.
 - 5.7 Copy of GST, PAN, TIN document
 - 5.8 Whether tenderer is manufacturer / accredited agent / sole representative, indicate details of principal's name & address. The offers of tenderer who are not manufacturer or direct authorized agent will be summarily rejected. Sub- distributors will not be accepted.
 - 5.9 Non – blacklisting certification that the firm has not been blacklisted in the past by any government/Private institution and certification for No Vigilance/CBI case pending against the firm/supplier by making an affidavit on non – judicial stamp paper of ₹100/-.
 - 5.10 Declaration towards acceptance of all terms and conditions should also be provided.
6. Financial Bid should comprise of-
- 6.1. Quotation should be only for DAP i.e. **NIMHANS Door Delivery Price inclusive of main equipment cost, each article wise/spares rates, taxes, other Government levies, Customs duty, any local agency commission, transportation, delivery of the equipment to the Institute premises, installation and commissioning, etc.** with a detailed break up mentioning manufacturers name, License number and name of the brand/make. Tender bids without price bid/quotation will be rejected. **CMC COST SHOULD BE QUOTED IN INR ONLY.**
Detailed breakup of the cost should be provided under icon “Action column” by clicking the % symbol which will be mentioned as Added statutory components.
 - 6.2. The tenderer should also quote **5 years CMC cost for the post warranty period in INR** under Action column in E procurement portal but CMC cost will not be considered for successful tenderer decision.
 - 6.2.1 In case of Import, Customs Duty will be considered only at 5.50% of the total Purchase order value and the required Customs clearance documents will be provided by NIMHANS on prior intimation (atleast 5 working days prior intimation) of shipment along with details and relevant documents from the supplier.
 - 6.2.2 If the tenderer is quoting in Indian Rupees (INR) for items NOT MANUFACTURING IN INDIA (NMIC), the CUSTOM DUTY EXEMPTION



CERTIFICATE WILL NOT BE ISSUED BY THE INSTITUTE. The Rate quoted should be inclusive of Custom duty & other incidental charges.

7. Successful tenderer decision will be made on the basis of total cost of the equipment (Inclusive of all miscellaneous charges as mentioned in Clause 6.1) + 5 years CMC cost for the post warranty period (i.e. for 6TH to 10th year).

8. The cost of the “NEUROSURGICAL OPERATING MICROSCOPE” will be freezed for 1 Year from the date of purchase order; however Institute reserves the right to procure/reject the purchase of equipment with the successful tenderer on repeat order basis within 1 year from the date of purchase order.

9. The tender bids (technical and price bid) should be typewritten; every correction in the tender should be initialed along with seal by the tenderer, failing which the tender will be rejected. All pages of the bid submitted must be signed along with seal and sequentially numbered by the tenderer.

10. Evaluation of Bids:-

The technical bid of the tenderer will be evaluated to determine whether

- a. They are complete with respect to specifications.
- b. They are free from computational errors.
- c. The requisite documents have been submitted and properly signed.

11. Tender Opening:

a. The Technical bids will be unlocked through E procurement portal in the Committee Room, Adjacent to Registrar Chamber, NIMHANS, Bengaluru on the date specified in presence of tenderers or their representatives who choose to attend.

The Tenderers' or representatives who are present shall submit authorization letter along with copy of Photo id proof and shall sign a register evidencing their attendance.

b. The Financial bid of the technically qualified tenderer/s only will be opened on a notified date.

12. Equipment and its accessories should be covered with minimum warranty period of 5 years for normal or regular wear & tear from the date of complete installation (Ready to use in all respects). In case of software's, the validity of the license key should be clearly mentioned and should have user define provision with option to switch over from one system to other system of the same kind within the validity period.

13. Supply of spares should be guaranteed for a minimum period of 10 years from the date of supply or from the date of cessation of production of the model for 10 years, whichever is later, at the rates prevailing against payment.

14. Any modification or revision of bids after submission will not be entertained under any circumstances. Conditions such as “subject to the availability of stocks”, supplies will be made as and when supplies received from the principles etc., will not be considered under any circumstances.



15. A tenderer having once given a tender bid shall not withdraw it after its acceptance/opening and if does, the EMD paid by the tenderer will be forfeited and the tenderer is liable to make good the loss sustained.
16. If required, the tenderer should demonstrate the quoted model of the equipment at the institute during the technical evaluation, failing which their bid/offer shall be rejected. The tenderer will be intimated that they should get ready for demonstration. No request for extending time for demonstration will be entertained. Failure to demonstrate, their offer will be rejected.
17. The tenderer should supply the circuit diagram and instruction manual of the tendered equipment/s at the time of supply of the equipment.
18. Necessary training / instructions on operation of the system should be given by the qualified engineers of the tenderer firm to NIMHANS technical staff/s at free of cost after completion of the installation.
19. The successful tenderer should immediately submit an acceptance letter duly signed and sealed for the rate/s and offers agreed by both the parties to the Head of the Institution within reasonable time on receipt of the Purchase Order (Agreement Specimen will be enclosed with Purchase order & Stamp duty to be paid by the tenderer). The successful tenderer should also furnish a Bank guarantee only from a Nationalized bank to the extent of 10% of the total purchase order value, valid for 60 days beyond the completion of the warranty period of the equipment, no split period bank guarantee will be entertained.
In the event of the successful tenderer failed to supply the item/execute the agreement/submit the Bank Guarantee the EMD deposited by them shall stands forfeited.

20. Payment terms:

- 21.1 Payment (NEFT / RTGS / Wire transfer / LC) will be made only after good working condition of the equipment certified by the end user. **NO ADVANCE PAYMENT WILL BE ENTERTAINED.**
- 21.2 In case of foreign payments made by Letter of Credit (LC) or wire transfer, bank incidental charges within India will be borne by the Institute and outside India should be borne by tenderer.
- 21.3 Any amendment or extension of LC sought by the tenderer thereafter should be borne by the tenderer.

21. Uptime Guarantee:

Penalty Clause for non-functioning of equipment in term of hardship to the patients and financial loss to institute: 95% up time of 365 days (24 hours a day) that is from the day of successful handing over of the whole complex. The company takes the responsibility for the functioning of all the components and equipment, including the third party items supplied and included in the project. The total downtime annually for any reason/involvement of any of the components cannot exceed 5% (all inclusive). Subsequently if downtime exceeds 5% of 365 days, 1% of PO Value will be levied as penalty for every 24 hours of downtime until 7 days from the day of breakdown. If downtime exceeds 7 days the penalty will be 2% of PO Value from the date on



E-Procurement Global Tender No. STR–D3/NIMHANS/2019-20/IND565 30.12.2019

- which the equipment broke down beyond 5% permissible downtime. In addition to this, warranty period will be extended at double the rate of the downtime period.
22. If, at any time, during the said period, the supplier reduce the said prices of such Materials/Equipment or sales such Materials/Equipment to any other person/organization/ Institution at a price lower than the chargeable, the company shall forthwith notify such reduction or sale to the Director, NIMHANS and the price payable for the Materials supplied after the date of coming into force of such reduction or sale shall stand correspondingly reduced.
 23. The losses to NIMHANS, Bengaluru, if any incurred on account of purchase made elsewhere by failure, neglect or refusal on the part of the tenderer to supply according to the terms of agreement will be recovered from them. If any article or things supplied by the tenderer have been partially or wholly used or consumed in the hospital and they are subsequently found to be in bad condition, unsound, inferior in quality or description, not in accordance with samples or otherwise faulty or unit for use, the wholesome of the contract price or price of such articles or things will be recovered from the tenderer. The tenderer will not be entitled for any payment whatsoever, for such articles for infringements of the stipulation of the conditions or for justifiable reasons the contract may be terminated by the Director and the tenderer shall be liable for losses sustained by the NIMHANS on the consequences of the termination which may be recovered from the EMD/Bank Guarantee or from their invoices due to them. In the event of such amount being insufficient, the balance will be recovered personally from the tenderer.
 24. Any corrections/changes in the tender will be uploaded as corrigendum in the NIMHANS and E procurement websites only.
 25. If the tender last/opening date falls on any general/government/institute holiday(s), then the successive dates will be postponed by equivalent days of holiday(s), however the time remains unchanged.
 26. The Director reserves the right of ordering/not ordering/cancelling/increase or decrease the quantity and to reject any or all tender quotations without assigning any reason. The decision of the Director, NIMHANS, Bengaluru, shall be final in all the controversies that may arise in the matter. Any dispute arising out of this will be subject to the jurisdiction of the Court in Bengaluru.
 27. Failure to adhere any of the above terms and conditions the bid(s) may be rejected and EMD may be forfeited.
 28. None of the terms and conditions of the supplier shall be applicable to the purchase contemplated hereunder, irrespective of it being attached to any documents to be provided to NIMHANS. Such exercise shall have no meaning and binding effect unless the same is accepted by NIMHANS in writing.

NOTE: Please keep checking the NIMHANS and E-Procurement websites regularly for any further updates.

Sd/-,
Director



DECLARATION
(TO BE GIVEN BY THE TENDERER)

Name of the Item : NEUROSURGICAL OPERATING MICROSCOPE
Name of the company :

To

The Director,
National Institute of Mental Health & Neurosciences
(Institute of National Importance)
Post Box No. 2900,
Hosur Road, Bengaluru – 560 029

Dear Sir,

1. I/We hereby submit my/our tender for the
2. I/We have made requisite payment against EMD as per the tender document vide reference No. & date, else my tender bid may be rejected.
3. I/We have gone through all terms and conditions of the tender documents before submitting the same.
4. I/We hereby agree to all the terms and conditions, stipulated by the NIMHANS, in this connection including delivery, warranty, penalty etc. Quotations for each group are being submitted and shall be considered on their face value.
5. I/We undertake to sign the contract/agreement, if required, within reasonable time from the date of issue of the letter of acceptance, failing which our/my security money deposited may be forfeited and our/my name may be removed from the list of suppliers at the NIMHANS, Bengaluru.

NOTE: ALL TERMS & CONDITIONS SUCH AS TAXES/LEVIES ETC, HAS BEEN INDICATED IN THE QUOTATIONS FAILING WHICH IT WILL BE PRESUMED THAT THE RATES ARE INCLUSIVE OF ALL TAXES/LEVIES AND OTHER TERMS AND CONDITIONS ARE ALSO AS PER YOUR REQUIREMENTS.

Yours faithfully,

Signature of Tenderer & seal



E-Procurement Global Tender No. STR–D3/NIMHANS/2019-20/IND565 30.12.2019

CHECK LIST
(TO BE FILLED BY THE TENDERER)

1	Name & Address of the Tenderer		
2	Name & Address Of the Manufacturer		
3	Name of the Equipment & Model Quoted	NEUROSURGICAL OPERATING MICROSCOPE	
4	Validity of the Quotation	Four months from the due date	
5	a. Delivery Period b. Warranty Period c. Installation Period	5 years	
6	Tender Bid details (Yes or No against each item)	a. Application Fee submitted	
		b. EMD submitted	
		c. Brochure/Catalogue uploaded	
		d. Technical Compliance Statement	
		e. Manufacturer Proprietary certificate uploaded	
		f. Pre requirements details uploaded	
		g. List of users uploaded	
		h. Copy of GST/PAN/TIN & Bank details uploaded	
		i. Distributor authorization letter uploaded	
		j. Non-blacklisting certification uploaded	
		k. Declaration enclosed	
		l. Equipment Door delivery cost quoted	
		m. AMC & CMC cost for 5 years post warranty period	
8	Training will be provided (Yes or No)		
9	a. Whether after sales, service is available in Bengaluru? If yes, quote		
	b. What is the arrangement for post contract / warranty monitoring of the equipment?		
10	Details of the EMD Submitted (Reference no., date & bank details)		
11	Any Other Information (Enclose separately in letter head – Yes or No)		

Signature of Tenderer along with seal



NEUROSURGICAL OPERATING MICROSCOPE

For Technical Clarifications contact:

Dr. Dwarakanath Srinivas, Prof & Head – Mobile: - +91-80-9480829605

Office Landline: +91-80-26995403/5724

Dept. of Neurosurgery, Faculty Building, NIMHANS, Hosur Road, Bengaluru

Email ID: dwarakaneuro@yahoo.com & surgery.nimhans@gmail.com

TENDER SPECIFICATIONS FOR NEUROSURGICAL OPERATING MICROSCOPE

SL No	Microscope	Description
1	Classification	Class 1
2	Optics Carrier	Please specify the height
3	Working distance	200- 500mm (± 25 mm) Variable through motorised multifocal lens, activated through Hand switch, Footswitch and through control panel. Manually adjustable override.
4	Magnification range	1.5x - 17x with minimum 10x or higher eyepiece
5	Field of View	12.5mm - 175mm with 10X Eyepiece
6	Focusing	Motorised, via multifocal lens activated through Hand Switch, Footswitch & Control panel. Manually adjustable override has to be available
7	Eyepieces	Wide - field eyepiece for spectacle wearers 10X or 12.5X, dioptric setting +/- 5 with adjustable eye cup
8	Observation	Wide range of observation: Two observers should view simultaneously.
9	Light Source	Main Lamp should be minimum 300W Xenon Illumination
10	Emergency Lamp	Back-up also should be 300 W Xenon
11	Illumination Power supply	Please mention the details of illumination
12	Illumination Field Diameter	Should have Built in automatic zoom-synchronized illumination field diameter, with manual override and reset feature
13	Automated illumination control	Should have Automated illumination Brightness control should be linked to working distance
14	Binocular tube	Flexible ergonomic height adjustment for optimal body position at the microscope.
15	Asepsis for all controls	Sterilisable Protective glass for objective. Sterilisable components for all drive knobs/drapes
16	Co-Observer attachment	Stereo Co-Observer attachment for Cranial and for Spine attachment
17	Rotation of optics	Minimum 500 degrees
18	Lateral Tilt	-45 to +45 or better
19	Inclination Tilt	-30 to +120 or better
20	Handles /Joystick	Controls for microscope zoom position, variable working distance via multifocal lens, XY movement, and the 6 electro-magnetic brakes
21	Balancing	Full Auto Balancing of the whole system by one button push / touch.
22	Inter-operative Auto balance	The System should preferably have intra-operative Auto balance facility during the case without breaking sterility



E-Procurement Global Tender No. STR–D3/NIMHANS/2019-20/IND565 30.12.2019

STAND		
23	Floor stand	Floor Stand with 6 Electro-magnetic brakes
24	Range of Cantilever	To be mentioned
25	Compact Base	Should be in the range of 650- 750 X 650-750
26	Control Unit	Latest advanced intuitive touch panel technology with built in diagnostic Menu and indication for Lamp remaining hours.
27	HD camera	HD 1080P or higher Camera with excellent video output. A 3-chip camera is preferable. Please mention if it is available and the quote should include the cost of a 3-Chip camera.
28	Recording System	High Definition (1080P). Recording system with built in Monitor Live duplicating - file storage on any of the external storage device viz., Pen drive, USB storage device etc.,
		Video format - MPEG -4
		HD input/output
		Minimum Storage capacity of 2TB HDD
		Compatible Intuitive Operating System
		Video Compression should be MPEG4 & Still Image bmp & jpeg
29	Navigation	Microscope should be compatible with all the existing major navigation systems in the market for image guided surgery. All the necessary software for the same should be incorporated. The microscope should be compatible and should have facility for integration with neuronavigation.
		The microscope should have facility for fluorescence guided tumor surgery and vascular surgery (for intra operative ICG angiography.)The tumor fluorescence should include attachment (both hardware and software) for either ALA or yellow fluorescence. The Department should be able to select either of the two. All the three, ICG, ALA and yellow fluorescence should be approved by the FDA, CE Mark and the appropriate Indian regulator It should be integrated with the microscope. Please mention the camera configuration and resolution available for the fluorescence surgery.
30	Fluorescence	
31	Image injection in the eyepiece	Please mention details
32		The equipment should have international certification like CE, IEC etc. Please provide the names of institutes/hospitals (Neurosurgery only) where the quoted model is installed.
33		The supplier should be of good standing in Indian market for at least 5 years. A demonstration of the quoted equipment should be possible if requested. Years. Warranty for 5 years.

WARRANTY: 5 YEARS

COMPREHENSIVE MAINTENANCE CONTRACT [CMC]: 6TH to 10TH Year (5 Years)