



NATIONAL INSTITUTE OF MENTAL HEALTH AND NEURO SCIENCES  
(Institute of National Importance), Bengaluru- 560 029

राष्ट्रीय मानसिक स्वास्थ्य और तंत्रिका विज्ञान संस्थान, (राष्ट्रीय महत्व संस्थान), बंगलूरु - 560 029  
ರಾಷ್ಟ್ರೀಯ ಮಾನಸಿಕ ಆರೋಗ್ಯ ಮತ್ತು ನರ ವಿಜ್ಞಾನ ಸಂಸ್ಥೆ, (ರಾಷ್ಟ್ರೀಯ ಪ್ರಾಮುಖ್ಯತಾ ಸಂಸ್ಥೆ), ಬೆಂಗಳೂರು - 560 029

Phone 26995023/5913/5923/5024/5025/5780 Website <http://www.nimhans.ac.in/tender>  
Fax 080-26571563/26564830/2121/6811 E-mail [aaos@nimhans.ac.in](mailto:aaos@nimhans.ac.in)



## PURCHASE SECTION

STR-D1/389/EQPT/LTE/ Polypropylene cages for Mice & Rats/CARF/2020-21

02.02.2021

# LIMITED TENDER ENQUIRY

**Sub: Limited Tender Enquiry (Single bid system) towards supply of “Polypropylene/polycarbonate cages for laboratory Mice and Rats”**

Dear Sir/Madam,

With reference to the subject cited above, kindly make arrangements to submit your quotation in a sealed cover for supply of the following item, as detailed below addressed to **“The Administrative Officer(S), Purchase Section, NIMHANS, Hosur Road, Bengaluru – 560 029”** before 4 pm on 03.03.2021 (due date). An early compliance is requested.

Downloading of Tender documents from website		<a href="http://www.nimhans.ac.in">www.nimhans.ac.in</a>
<i>End User Contact Details</i>		
Name	Dr. B. S. Shankaranarayana Rao	
Department	Central Animal Research Facility (CARF)	
Designation	Professor, Officer in charge and Registrar	
Office No.	+91 - 80 – 2699 5175/5005/5006/ 5040/5039	
Email	<a href="mailto:bssrao.nimhans@gmail.com">bssrao.nimhans@gmail.com</a> <a href="mailto:carf@nimhans.ac.in">carf@nimhans.ac.in</a>	
SL. NO.	<b>SPECIFICATIONS</b>	
01.	<b>Specifications for Polypropylene/polycarbonate cages for laboratory mice.</b>	<b>Quantity required</b>
	<b>A. <u>Polypropylene mice cages:</u></b> <ol style="list-style-type: none"><li>Should be of Dimensions: 29X22X14cm (Length x Width x Height) and suitable for housing of laboratory mice under conventional system</li><li>All mice cages should fit into hanging type SS trolley racks</li><li>Cages should be Autoclavable at 121<sup>0</sup> C temperature and pressure 15 LB/ Sq. inch for 15 minutes</li><li>Should be Food Grade quality Polypropylene material with uniform thickness</li><li>Excellent stack ability for ease of handling and storage</li><li>Top grill should be made of S.S. 304 material with duly spot welded and finished joints</li><li>AISI 304 quality stainless steel material</li><li>Provision for fixing water bottle</li><li>Provision for Feed Pallets</li><li>Fitted with two hooks of brass material for locking the grill to</li></ol>	<b>200 Nos.</b>



## PURCHASE SECTION

	<p>the top of the body of cage</p> <ol style="list-style-type: none"><li>11. Fitted with tag holder</li><li>12. Gap between S.S. rods should be around 6mm or less</li><li>13. Should ensure perfect fit and locking throughout the complete cage body</li><li>14. Grill tops should be designed to fit in second grill for ease of stacking with foldable partition for food and water bottle compartment</li><li>15. Test certificate of S.S. 304 quality for grill material must be supplied along with cage.</li><li>16. Should be supplied with autoclave resistant card holders in sufficient numbers</li></ol> <p><b>B. Polycarbonate cages for mice:</b></p> <ol style="list-style-type: none"><li>1. Should be of Dimensions: 29X22X14cm (Length x Width x Height) and suitable for housing of laboratory mice under conventional system</li><li>2. All mice cages should fit into hanging type SS trolley racks</li><li>3. Cages should be Autoclavable at 121<sup>0</sup> C temperature and pressure 15 LB/ Sq. inch for 15 minutes</li><li>4. Fully transparent body</li><li>5. Food Grade quality material Polycarbonate material with uniform thickness</li><li>6. Excellent stack ability for ease of handling and storage</li><li>7. Grill tops must be made of S.S. 304 material with duly spot welded and finished joints</li><li>8. AISI 304 quality stainless steel material</li><li>9. Provision for fixing water bottle</li><li>10. Provision for Feed Pallets</li><li>11. Fitted with two hooks of brass material for locking the grill to the top of body of cage</li><li>12. Fitted with tag holder</li><li>13. Gap between S.S. rods should be 6 mm or less</li><li>14. Should ensure perfect fit and locking throughout the complete cage body</li><li>15. Grill tops should be designed to fit in second grill for ease of stacking with foldable</li><li>16. Partition for food and water bottle compartment</li><li>17. Test certificate of S.S. 304 quality for grill material must be supplied along with cage.</li><li>18. Should be supplied with autoclave resistant card holders in sufficient numbers</li></ol>	
	<p><b>Specifications for Polypropylene/polycarbonate cages for laboratory Rats.</b></p>	



## PURCHASE SECTION

02.	<p><b>A. <u>Polypropylene rat cages:</u></b></p> <ol style="list-style-type: none"><li>1. Should be of Dimensions: 43x29x15cm (Length x Width x Height) and suitable for housing of laboratory rats under conventional system</li><li>2. All rat cages should fit into hanging type SS trolley racks</li><li>3. Cages should be autoclavable at 121<sup>0</sup> C temperature and pressure 15 LB/ Sq. inch for 15 minutes</li><li>4. Should be Food Grade quality Polypropylene material with uniform thickness</li><li>5. Excellent stack ability for ease of handling and storage</li><li>6. Top grill should be made of S.S. 304 material with duly spot welded and finished joints</li><li>7. AISI 304 quality stainless steel material</li><li>8. Provision for fixing water bottle</li><li>9. Provision for Feed Pallets</li><li>10. Fitted with two hooks of brass material for locking the grill to the top of the body of cage</li><li>11. Fitted with tag holder</li><li>12. Gap between S.S. rods should be around 7mm or less</li><li>13. Should ensure perfect fit and locking throughout the complete cage body</li><li>14. Grill tops should be designed to fit in second grill for ease of stacking with foldable partition for food and water bottle compartment</li><li>15. Test certificate of S.S. 304 quality for grill material must be supplied along with cage.</li><li>16. Should be supplied with autoclave resistant card holders in sufficient numbers</li></ol> <p><b>B. <u>Polycarbonate cages for mice:</u></b></p> <ol style="list-style-type: none"><li>1. Should be of Dimensions: 43x29x15cm (Length x Width x Height) and suitable for housing of laboratory rats under conventional system</li><li>2. All rat cages should fit into hanging type SS trolley racks</li><li>3. Cages should be Autoclavable at 121<sup>0</sup> C temperature and pressure 15 LB/ Sq. inch for 15 minutes</li><li>4. Fully transparent body</li><li>5. Food Grade quality material Polycarbonate material with uniform thickness</li><li>6. Excellent stack ability for ease of handling and storage</li><li>7. Grill tops must be made of S.S. 304 material with duly spot welded and finished joints</li><li>8. AISI 304 quality stainless steel material</li><li>9. Provision for fixing water bottle</li><li>10. Provision for Feed Pallets</li></ol>	300 Nos.
-----	---	----------



NATIONAL INSTITUTE OF MENTAL HEALTH AND NEURO SCIENCES  
(Institute of National Importance), Bengaluru- 560 029

राष्ट्रीय मानसिक स्वास्थ्य और तंत्रिका विज्ञान संस्थान, (राष्ट्रीय महत्व संस्थान), बंगलूरु - 560 029  
ರಾಷ್ಟ್ರೀಯ ಮಾನಸಿಕ ಆರೋಗ್ಯ ಮತ್ತು ನರ ವಿಜ್ಞಾನ ಸಂಸ್ಥೆ, (ರಾಷ್ಟ್ರೀಯ ಪ್ರಾಮುಖ್ಯತಾ ಸಂಸ್ಥೆ), ಬೆಂಗಳೂರು - 560 029

Phone 26995023/5913/5923/5024/5025/5780 Website <http://www.nimhans.ac.in/tender>  
Fax 080-26571563/26564830/2121/6811 E-mail [aaos@nimhans.ac.in](mailto:aaos@nimhans.ac.in)



## PURCHASE SECTION

11.	Fitted with two hooks of brass material for locking the grill to the top of body of cage
12.	Fitted with tag holder
13.	Gap between S.S. rods should be 7 mm or less
14.	Should ensure perfect fit and locking throughout the complete cage body
15.	Grill tops should be designed to fit in second grill for ease of stacking with foldable
16.	Partition for food and water bottle compartment
17.	Test certificate of S.S. 304 quality for grill material must be supplied along with cage.
18.	Should be supplied with autoclave resistant card holders in sufficient numbers

Sd/-  
Registrar

### Terms & Conditions

- Sealed quotation (single bid)** in has to be submitted only by either manufacturer of the articles or the authorized local supplier of the manufacturer.
- Validity of the bid should be for 120 days from the due date.
- Enquiry reference number along with date and due date should be superscribed on outer cover.
- Quantity of items may be increased / decreased as per requirement.
- Financial bid should be in INR for NIMHANS door delivery price** and total cost with detailed breakup of cost towards Item wise price, taxes, any surcharges (such as transportation, installation, customs duty, etc.), warranty/expiry period, delivery period, contact person details (such as Mobile No., E-mail, local supplier address) should be clearly mentioned.
- Quotation must be submitted along with relevant documents counter signed along with seal by the bidder (such as Authorization certificates, Purchase order related to previous supply to Institutes/Hospitals, Catalogues, Data Sheet, PAN number, GST Registration copy, Bank details, Technical Compliance Statement, Proprietary certificate from the manufacturer (if applicable), Pre requirements required for using chemicals, Non – blacklisting certification that the firm has not been blacklisted in the past by any Government / Private institution and certification for No Vigilance/CBI case pending against the firm/supplier by making an affidavit on non – judicial stamp paper of `100/-, Declaration, etc.).** If not submitted quotation will be treated as incomplete and may be rejected.
- If the purchase of article is approved by the Institute, the article should be supplied/installed within the delivery period quoted else penalty will be levied as per institute rules.
- Equipment and its accessories should be covered with **minimum warranty period of 5 years** for normal or regular wear & tear from the date of complete installation (Ready to use in all respects).
- The cost of the “Polypropylene/polycarbonate cages for laboratory Mice and Rats” will be freezed for 1 Year from the date of purchase order; however Institute reserves the right to procure/reject the purchase of equipment with the successful tenderer on repeat order basis within 1 year from the date of purchase order.**



NATIONAL INSTITUTE OF MENTAL HEALTH AND NEURO SCIENCES  
(Institute of National Importance), Bengaluru- 560 029

राष्ट्रीय मानसिक स्वास्थ्य और तंत्रिका विज्ञान संस्थान, (राष्ट्रीय महत्व संस्थान), बंगलूरु - 560 029  
ರಾಷ್ಟ್ರೀಯ ಮಾನಸಿಕ ಆರೋಗ್ಯ ಮತ್ತು ನರ ವಿಜ್ಞಾನ ಸಂಸ್ಥೆ, (ರಾಷ್ಟ್ರೀಯ ಪ್ರಾಮುಖ್ಯತಾ ಸಂಸ್ಥೆ), ಬೆಂಗಳೂರು - 560 029

Phone 26995023/5913/5923/5024/5025/5780 Website <http://www.nimhans.ac.in/tender>  
Fax 080-26571563/26564830/2121/6811 E-mail [aaos@nimhans.ac.in](mailto:aaos@nimhans.ac.in)



## PURCHASE SECTION

11. Sealed quotation shall be addressed to  
**THE ADMINISTRATIVE OFFICER(S),  
PURCHASE SECTION, NIMHANS,  
HOSUR ROAD, BENGALURU – 560 029**
10. **Quotation received after last date and time will not be considered under any circumstances.**
11. However, the Institute reserves the right to accept or reject any/all quotations without assigning any reason whatsoever.
12. **NO ADVANCE PAYMENT.** Bill will be passed for payment to accounts section only after satisfactory supply/installation certification of the articles from the concerned Head of the Department/Section. Further payment will be made by accounts section.
13. If necessary the firm may be called for presentation/demo of the articles.
14. For equipments worth ₹ 2,00,000/- and above, Agreement has to be executed on ₹ 200/- stamp paper (Specimen will be provided with purchase order) and Bank Guarantee has to be executed from a Nationalized Bank for 10% of total value valid for a period of 60 days beyond the warranty period, else payment will not be processed .
15. None of the terms and conditions of the supplier shall be applicable to the purchase contemplated hereunder, irrespective of it being attached to any documents to be provided to NIMHANS. Such exercise shall have no meaning and binding effect unless the same is accepted by NIMHANS in writing.

